

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) - Italia - Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: CARDIOLIPIN-G - CHORUS

Code	86046		
Lot No.	627	2021-04	2020-04

REAGENTS

Label	Description	Lot.	Exp	No
D3311	CARDIOLIPIN-G	627	2021-04	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	627	2021-04	1
C3311	CONTROL SERUM CARDIOLIPIN-G	627	2021-04	1

CONTROLS

Found results	Expected results
Calibrator (mOD)	250
Control Serum (GPLU/mL)	68,9
	≥ 100
	42,9-89,5

CALIBRATOR


3	3	5	6	2	7	6	3	1	9	3	3	2	8	5	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


3	3	5	6	2	7	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.