

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SI) - Italia - Via delle Rose 10
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: SANITIZING SOLUTION

Code	83604		
Lot No.	067	Expiry	2024-09

REAGENTS

Label	Description	Lot.	Exp	No
B2563	SANITIZING SOLUTION	067	2024-09	1

CONTROLS

	Found results	Expected results
Density	1.062	1.060-1.100
pH	13,4	12-14

The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi *Simone Baiocchi*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE: 08.10.2021

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

na