

**DIESSE DIAGNOSTICA SENESE SpA**  
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
 20144 Milano (Italy) Via A. Solari 19 sc. 6  
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

**QUALITY CONTROL CERTIFICATE**

**KIT NAME: CHORUS ECHINOCOCCUS IgG**

|                |           |
|----------------|-----------|
| <b>Code</b>    | 81284     |
| <b>Lot No.</b> | 863       |
|                | 🕒 2023-08 |
|                | 📄 2022-08 |

**REAGENTS**

| Label | Description                    | Lot. | Exp     | No  |
|-------|--------------------------------|------|---------|-----|
| D0010 | Echinococcus IgG               | 863  | 2023-08 | 2X6 |
| B0010 | CALIBRATOR Echinococcus IgG    | 863  | 2023-08 | 1   |
| C0010 | CONTROL SERUM Echinococcus IgG | 863  | 2023-08 | 1   |


**CONTROLS**

|                       | Found results | Expected result |
|-----------------------|---------------|-----------------|
| CALIBRATOR (mOD)      | 560           | ≥100            |
| CONTROL SERUM (INDEX) | 1,5           | 1,2-2,2         |

CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 3 | 5 | 8 | 6 | 3 | 6 | 0 | 0 | 0 | 0 | 0 | 6 | 0 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader **YES**  Operator Signature ..... *SB* .....



CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 3 | 5 | 8 | 6 | 3 | 6 | 1 | 2 | 2 | 2 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader **YES**  Operator Signature ..... *SB* .....



The test has been carried out in accordance with the package insert

TESTED BY: Sara Benocci *SB*  
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alioudi*  
 DATE *08.09.22*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.