DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgG AVIDITY - CHORUS

Code	81098				
Lot No.	914	⌛	2024-02		2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2531	TOXOPLASMA IgG AVIDITY	914	2024-02	2x6
C2531	CONTROL SERUM HIGH TOXOPLASMA IgG AVIDITY	914	2024-02	1
G2531	SAMPLE DILUENT	914	2024-06	1

CONTROLS

	Found results	Expected results
Control Serum (%)	56,04 %	> 50 ≤ 90 %

CALIBRATOR

Label checked with barcode reader YES ☐ Operator Signature .....

STICK LABEL

CONTROL SERUM

0	8	3	9	1	4	6	5	0	9	0	0	0	0	0	0	0	1

Label checked with barcode reader YES ☒ Operator Signature .....RB.....



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 24.03.'23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.