

DIESSÉ DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code	81270				
Lot No.	921		2024-03		2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	921	2024-03	6X6
B9660	CALIBRATORE Tetanus IgG	921	2024-03	1
C9660	CONTROL SERUM Tetanus IgG	921	2024-03	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	526	≥ 100
Control (mIU/mL)	407,9	346,2-750,0

CALIBRATOR

9	0	2	9	2	1	6	3	2	3	3	5	4	6	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

902921632336465000  
Tetanus IgG  
CALIBRATOR 0.176 mL IVD  
LOT 921  
2024-03 2°C  
DIESSÉ Monteriggioni Italy CE

CONTROL SERUM

9	0	2	9	2	1	6	1	5	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

902921615290000001  
Tetanus IgG  
CONTROL+ 0.426 mL IVD  
LOT 921  
2024-03 2°C  
DIESSÉ Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents...

DATE 06.04.2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.