

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgM -CHORUS

Code	81021		
Lot No.	926	2024-04	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2221	Herpes simplex 1+2 IgM	926	2024-04	6x6
B2221	CALIBRATORE Herpes simplex 1+2 IgM	926	2024-04	1
C2221	CONTROL SERUM Herpes simplex 1+2 IgM	926	2024-04	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	466	≥ 100
Control Serum (INDEX)	2,2	1,8-3,4
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

1	5	6	9	2	6	6	0	0	0	0	0	3	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

15692660000390000
Herpes simplex 1+2 IgM
CALIBRATOR 0.175 mL IVD
LOT 926
2024-04 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	5	6	9	2	6	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

15692661834000001
Herpes simplex 1+2 IgM
CONTROL+ 0.425 mL IVD
LOT 926
2024-04 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 05/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.