



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgG -CHORUS

Code	81020				
Lot No.	934		2024-08		2023-05

REAGENTS

Label	Description	Lot.	Exp.	NO
D2211	Herpes simplex 1+2 IgG	934	2024-08	6X6
B2212	CALIBRATORE Herpes simplex 1+2 IgG	934	2024-08	1
C2211	CONTROL SERUM Herpes simplex 1+2 IgG	934	2024-08	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	373	≥ 100
Control Serum (INDEX)	3,1	2,0-3,7
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0	0	4	9	3	4	7	0	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

0	0	4	9	3	4	7	2	0	3	7	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 11/05/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.