

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

### QUALITY CONTROL CERTIFICATE

**KIT NAME** SANITIZING SOLUTION

Code	83604		
Lot No.	072-A	2026-03	2023-03

### REAGENTS

Label	Description	Lot.	Exp.	NO
B2563	SANITIZING SOLUTION	072	2026-03	1

### CONTROLS

	Found results	Expected results
Density	1.061	1.060-1.100
pH	13,6	12-14

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 15.05.23