

QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr VCA IgM II - CHORUS

Code	81054		
Lot No.	928	2024-07	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D9490	Epstein-Barr VCA IgM II	928	2024-07	6x6
B9490	CALIBRATOR Epstein-Barr VCA IgM II	928	2024-07	1
C9490	CONTROL SERUM Epstein-Barr VCA IgM II	928	2024-07	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	520	≥ 100
Control Serum (INDEX)	2,1	1,4-2,7
Negative Control (INDEX)	0,0	< 0,7

CALIBRATOR

0	3	0	9	2	8	7	0	0	0	0	0	0	4	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*



CONTROL SERUM

0	3	0	9	2	8	7	1	4	2	7	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *23.05.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.