



QUALITY CONTROL CERTIFICATE

KIT NAME WW DAT Brucella Positive Control

Code	26007				
Lot. No	172		2024-07		2023-04

REAGENTS


CODE	LABEL	DESCRIPTION	Lot No.	EXP	NO
PF21250/BR	B4101	WWDAT Brucella Positive Control	172	2024-07	1

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
Brucella	1/320	≥ 1/40

The test has been carried out in accordance with the package insert

TESTED BY: Katia Tanzini 

APPROVED BY: (RESPONSIBLE HEAD, Quality Control Reagents) 

DATE 20.06.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.