

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME SANITIZING SOLUTION

Code	83604		
Lot No.	072-B	2026-03	2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
B2563	SANITIZING SOLUTION	072	2026-03	1

CONTROLS

	Found results	Expected results
Density	1.061	1.060-1.100
pH	13,6	12-14

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE *10.06.2023*