

DIESSE DIAGNOSTICA SENESE SpA
 53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
 20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: PARVOVIRUS B19 IgM - CHORUS

| | | | |
|----------------|-------|---------|---------|
| Code | 81152 | | |
| Lot No. | 943 | 2024-06 | 2023-06 |

REAGENTS

| Label | Description | Lot. | Exp | No |
|-------|----------------------------------|------|---------|-----|
| D0250 | Parvovirus B19 IgM | 943 | 2024-06 | 6X6 |
| B0250 | CALIBRATORE Parvovirus B19 IgM | 943 | 2024-06 | 1 |
| C0250 | CONTROL SERUM Parvovirus B19 IgM | 943 | 2024-06 | 1 |


CONTROLS

| | Found results | Expected result |
|-----------------------|---------------|-----------------|
| CALIBRATOR (mOD) | 320 | ≥100 |
| CONTROL SERUM (INDEX) | 2,3 | 1,9-3,6 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 2 | 2 | 9 | 4 | 3 | 6 | 0 | 0 | 0 | 0 | 0 | 3 | 7 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES Operator Signature *HB*



CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 2 | 2 | 9 | 4 | 3 | 6 | 1 | 9 | 3 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *HB*



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo *Antonella Leo*
 APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*
 DATE *23/06/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.