



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code	86014		
Lot No.	945	2024-06	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	945	2024-06	6X6
B3141	CALIBRATORE ANA-Screen	945	2024-06	1
C3141	CONTROL SERUM ANA-Screen	945	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	373	≥ 100
Control Serum (Ndx)	3,3	2,5-4,3

CALIBRATOR

1	1	5	9	4	5	6	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

11694660000300000
ANA-Screen
CALIBRATOR 0.178 mL IVD
LOT 945
2024-06 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	1	5	9	4	5	6	2	3	4	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

116946623430000001
ANA-Screen
CONTROL+ 0.428 mL IVD
LOT 945
2024-06 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

Aliaudi

DATE *10/07/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.