



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgG -CHORUS

Code	81060
Lot No.	946
	2024-06
	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	946	2024-06	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	946	2024-06	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	946	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	362	≥ 100
Control Serum (UA/mL)	76,2	59,4-144,4
Negative Control (UA/mL)	<5,0	< 5

CALIBRATOR

0 0 8 9 4 6 6 2 2 0 3 6 3 0 5 4 0 0

Label checked with barcode reader YES Operator Signature *FB*

008946622036306400
Helicobacter pylori IgG
CALIBRATOR 0.425 mL IVD
LOT 946
2024-06
2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

0 0 8 9 4 6 6 2 3 4 4 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *FB*

008946623440000001
Helicobacter pylori IgG
CONTROL 0.425 mL IVD
LOT 946
2024-06
2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *10.07.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.