

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME SM - CHORUS

Code 86016

Lot No. 945



2024-09



2023-06

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3161	SM	945	2024-09	6X6
B3161	CALIBRATORE SM	945	2024-09	1
C3161	CONTROL SERUM SM	945	2024-09	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	770	$\geq 100$
Control Serum (AU/mL)	35,1	23,3-65,0

### CALIBRATOR


1	9	2	9	4	5	7	2	1	9	2	6	4	8	3	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



### CONTROL SERUM


1	9	2	9	4	5	7	1	5	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 20/07/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.