



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME LKM-1 CHORUS

Code	86080		
Lot No.	943	2024-06	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3641	LKM-1	943	2024-06	6X6
B3641	CALIBRATORE LKM-1	943	2024-07	1
C3641	CONTROL SERUM LKM-1	943	2024-07	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	640	≥ 100
Control Serum AU/mL	73,8	42,2-92,9

CALIBRATOR

7	0	3	9	4	3	6	4	1	9	3	3	4	2	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature HB



CONTROL SERUM

7	0	3	9	4	3	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature HB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... HB

DATE 24/07/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.