



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI IgA-CHORUS**

Code	81062			
Lot No.	944	2024-06	2023-07	

REAGENTS

Label	Description	Lot.	Exp.	NO
D2401	Helicobacter Pylori IgA	944	2024-06	6X6
B2401	CALIBRATORE Helicobacter Pylori IgA	944	2024-06	1
C2401	CONTROL SERUM Helicobacter Pylori IgA	944	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	308	≥ 100
Control Serum (INDEX)	3,9	2,3-4,3
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0 2 7 9 4 4 6 0 0 0 0 0 3 1 0 0 0 0

Label checked with barcode reader YES Operator Signature *MB*

02794460000310000
Helicobacter pylori IgA
CALIBRATOR 0,178 mL IVD
LOT 844
2024-08 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

0 2 7 9 4 4 6 2 3 4 3 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *MB*

02794462343000001
Helicobacter pylori IgA
CONTROL+ 0,428 mL IVD
LOT 844
2024-08 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *MB*

DATE *26/07/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.