



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code	86074/12		
Lot No.	945	2024-06	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	945	2024-06	2x6
B3591	CALIBRATORE a-TPO	945	2024-06	1
C3591	CONTROL SERUM a-TPO	945	2024-06	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	360	≥ 100
Control Serum (IU/mL)	436,1	233,3-533,3

CALIBRATOR

3	0	1	9	4	5	6	5	1	9	3	7	3	4	5	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature Beni



CONTROL SERUM

3	0	1	9	4	5	6	2	1	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature Beni



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Beni

DATE 26.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.