



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME **88000-F79 GLUTEN**

Code	88000-F79		
Lot No.	943	2024-06	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D5011	88000-F79 GLUTEN	943	2024-06	2X6
B5010	CALIBRATORE 88000-F79 GLUTEN	943	2024-06	1
C5010	CONTROL SERUM 88000-F79 GLUTEN	943	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	290	≥ 100
Control Serum (KU/L)	17,2	10,0-41,5

CALIBRATOR

6	3	2	9	4	3	6	5	2	1	6	6	1	7	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature ..... *SB*

632943652166178100  
Gluten  
CALIBRATOR 0.250 mL [IVD]  
LOT 943  
2024-06 2°C 8°C  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6	3	2	9	4	3	6	4	1	7	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature ..... *SB*

63294364177000001  
Gluten  
CONTROL 0.480 mL [IVD]  
LOT 943  
2024-06 2°C 8°C  
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *28/10/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.