

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-M - CHORUS

Code	86052		
Lot No.	945	2024-06	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3371	BETA 2-GLYCOPROTEIN-M	945	2024-06	6X6
B3371	CALIBRATORE BETA 2-GLYCOPROTEIN-M	945	2024-06	1
C3371	CONTROL SERUM BETA 2-GLYCOPROTEIN-M	945	2024-06	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	400	> 100
Control Serum (AU/mL)	61,7	38,3-88,9

CALIBRATOR

2	0	5	9	4	5	6	5	3	1	4	9	2	9	6	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

206945653149296400
Beta-2-Glycoprotein-M
CALIBRATOR 0.176 mL [IVD]
LOT 945
2024-06
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	0	5	9	4	5	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

206945624460000001
Beta-2-Glycoprotein-M
CONTROL 0.425 mL [IVD]
LOT 945
2024-06
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 22.08.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.