



DIESSE DIAGNOSTICA SENESE Spa
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-H2 HOUSE DUST HOLLISTER-STIER LABS

Code	88000-H2		
Lot No.	963	2024-08	2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D3361	88000-H2 HOUSE DUST HOLLISTER-STIER LABS	963	2024-08	2X6
B3360	CALIBRATORE 88000-H2 HOUSE DUST HOLLISTER-STIER LABS	963	2024-10	1
C3360	CONTROL SERUM 88000-H2 HOUSE DUST HOLLISTER-STIER LABS	963	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR

6	4	1	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature 88



CONTROL SERUM

6	4	1	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature 88



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Busa

DATE 31.08.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.