

QUALITY CONTROL CERTIFICATE

KIT NAME TOSCANA VIRUS IgM - CHORUS

Code	81068		
Lot No.	947	2024-06	2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D2422	TOSCANA VIRUS IgM	947	2024-06	6X6
B2422	CALIBRATORE TOSCANA VIRUS IgM	947	2024-07	2
C2422	CONTROL SERUM TOSCANA VIRUS IgM	947	2024-07	2

CONTROLS

	Found results	Expected results
Calibrator (mOD)	440	≥ 100
Control Serum (Index)	2,5	1,7-3,2
Negative Control (Index)	0,1	<0,7

CALIBRATOR

0	2	5	9	4	7	6	0	0	0	0	0	4	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

0	2	5	9	4	7	6	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE 31/08/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.