

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-M - CHORUS

Code	86052		
Lot No.	946	2024-06	2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D3371	BETA 2-GLYCOPROTEIN-M	946	2024-06	6X6
B3371	CALIBRATORE BETA 2-GLYCOPROTEIN-M	946	2024-06	1
C3371	CONTROL SERUM BETA 2-GLYCOPROTEIN-M	946	2024-06	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	425	≥ 100
Control Serum (AU/mL)	56,8	38,3-88,9

CALIBRATOR

2	0	5	9	4	6	6	5	3	1	4	9	2	9	6	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature



CONTROL SERUM

2	0	5	9	4	6	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 08/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.