

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

KIT NAME CHORUS PHOSPHOLIPID SCREEN IgG

Code	86106		
Lot No.	970	2024-09	2023-09

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D4010	Phospholipid Screen IgG	970	2024-09	6X6
B4010	CALIBRATORE Phospholipid Screen IgG	970	2024-09	1
C4010	CONTROL SERUM Phospholipid Screen IgG	970	2024-09	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	430	≥ 100
Control (AU/ml)	38,1	27,3-52,9

**CALIBRATOR**

9	7	4	9	7	0	6	6	1	4	2	5	3	8	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

974970661426386300  
Phospholipid Screen-G  
CALIBRATOR 0.450 mL IVD  
LOT 970  
2024-09 2°C -8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

9	7	4	9	7	0	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

974970618340000001  
Phospholipid Screen-G  
CONTROL+ 0.900 mL IVD  
LOT 970  
2024-09 2°C -8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FB*

DATE *18.09.2023*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.