

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** GLIADIN-G - CHORUS

Code	86056
Lot No.	970
	🕒 2024-09
	📦 2023-09

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	970	2024-09	6X6
B3411	CALIBRATORE Gliadin-G	970	2024-09	1
C3411	CONTROL SERUM Gliadin-G	970	2024-09	1


**CONTROLS**

	Found results	Expected results
CALIBRATOR (mOD)	530	≥ 100
CONTROL SERUM (AU/ml)	84,2	38,3-146,2

**CALIBRATOR**

5	0	2	9	7	0	6	3	1	9	2	6	4	5	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*



**CONTROL SERUM**

5	0	2	9	7	0	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*



**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FB*

DATE *18/09/2023*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.