

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** Vitamin D Total- CHORUS

Code	86900		
Lot No.	952	2024-07	2023-09

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	952	2024-07	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	952	2025-05	1
C9703	CONTROL SERUM 25OH Vitamin D Total	952	2025-05	4

**CONTROLS**

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	460	≥ 100
CONTROL SERUM (ng/ml)	33,7	29,2-37,8

**CALIBRATOR**

8	3	5	9	5	2	6	3	3	5	4	0	5	3	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

835952633540536000  
25OH Vitamin D Total  
CALIBRATOR 0.300 mL IVD  
LOT 952  
2025-05 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123 B9703

**CONTROL SERUM**

8	3	5	9	5	2	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

835952613250000001  
25OH Vitamin D Total  
CONTROL+ 0.300 mL IVD  
LOT 952  
2025-05 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123 C9703

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 19.09.23

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.