

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgM - CHORUS

Code	81041		
Lot No.	962	2024-08	2023-08

REAGENTS

Label	Description	Lot.	Exp.	NO
D2281	TOXOPLASMA IgM	962	2024-08	6X6
B2281	CALIBRATORE TOXOPLASMA IgM	962	2024-08	1
C2281	CONTROL SERUM TOXOPLASMA IgM	962	2024-08	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	440	≥ 100
Control Serum (INDEX)	2,0	1,4-2,6
Negative Control (INDEX)	0,1	<0,7

CALIBRATOR

1	3	1	9	6	2	6	0	0	0	0	0	5	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature RB



CONTROL SERUM

1	3	1	9	6	2	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature RB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... Agua

DATE 20/09/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.