



**QUALITY CONTROL CERTIFICATE**

**KIT NAME** GLIADIN-A - CHORUS

Code	86054		
Lot No.	946	2024-06	2023-06

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3391	Gliadin-A	946	2024-06	6X6
B3392	CALIBRATORE Gliadin-A	946	2024-06	1
C3392	CONTROL SERUM Gliadin-A	946	2024-06	1

**CONTROLS**

	Found results	Expected results
CALIBRATOR (mOD)	300	≥ 100
CONTROL SERUM (AU/ml)	83,4	42,0-130,0

**CALIBRATOR**

5	0	1	9	4	6	6	5	3	2	4	7	2	6	6	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

501946653247266700  
Gliadin-A  
CALIBRATOR 0.175 mL IVD  
LOT 946  
2024-06 2°C-8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

5	0	1	9	4	6	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

501946627500000001  
Gliadin-A  
CONTROL 0.425 mL IVD  
LOT 946  
2024-06 2°C-8°C  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 20/09/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.