



**QUALITY CONTROL CERTIFICATE**

**KIT NAME** 88000-M2 CLADOSPORIUM HERBARUM

Code	88000-M2		
Lot No.	963	2024-08	2023-09

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D1171	88000-M2 CLADOSPORIUM HERBARUM	963	2024-08	2X6
B1170	CALIBRATORE 88000-M2 CLADOSPORIUM HERBARUM	963	2024-10	1
C1170	CONTROL SERUM 88000-M2 CLADOSPORIUM HERBARUM	963	2024-10	1

**CONTROLS**

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

**CALIBRATOR**

6	2	5	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

625963651062248700  
Cladosporium herbarum  
CALIBRATOR 0.250 mL IVD  
LOT 963  
2024-10 2°C-8°C  
DIESE Monteriggioni Italy CE

**CONTROL SERUM**

6	2	5	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

62596363055000001  
Cladosporium herbarum  
CONTROL+ 0.450 mL IVD  
LOT 963  
2024-10 2°C-8°C  
DIESE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Anna*

DATE 27/09/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.