



**QUALITY CONTROL CERTIFICATE**

**KIT NAME** 88000- E5- DOG DANDER

Code	88000-E5		
Lot No.	963	2024-08	2023-09

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D6911	88000- E5 DOG DANDER	963	2024-08	2x6
B6910	CALIBRATORE 88000- E5 DOG DANDER	963	2024-10	1
C6910	CONTROL SERUM 88000- E5 DOG DANDER	963	2024-10	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

**CALIBRATOR**

6	4	8	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

648963651862248700  
Dog dander  
CALIBRATOR 0.250 mL IVD  
LOT 963  
2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

6	4	8	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

64896363050000001  
Dog dander  
CONTROL 0.450 mL IVD  
LOT 963  
2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *FB*

DATE 28/09/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.