



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code	81078		
Lot No.	970	2024-09	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D247I	VARICELLA IgG	970	2024-09	6x6
B247Z	CALIBRATORE VARICELLA IgG	970	2024-09	i
C247I	CONTROL SERUM VARICELLA IgG	970	2024-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	450	≥ 100
Control Serum (Index)	3,0	1,9-3,4
Negative Control (Index)	0,1	< 0,7

CALIBRATOR

0	1	4	9	7	0	6	0	0	0	0	0	3	8	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*

01497060000380000
Varicella IgG
CALIBRATOR 0.428 mL IVD
LOT 970
2024-09 2°C-8°C
DIESSE Monteriggioni Italy CE
B247Z

CONTROL SERUM

0	1	4	9	7	0	6	1	9	3	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *SB*

01497061934000001
Varicella IgG
CONTROL+ 0.428 mL IVD
LOT 970
2024-09 2°C-8°C
DIESSE Monteriggioni Italy CE
C247I

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *Blue*

DATE *28.09.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.