

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-M3 ASPERGILLUS FUMIGATUS

Code	88000-M3		
Lot No.	963	2024-08	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D0431	88000-M3 ASPERGILLUS FUMIGATUS	963	2024-08	2X6
B0430	CALIBRATORE 88000-M3 ASPERGILLUS FUMIGATUS	963	2024-10	1
C0430	CONTROL SERUM 88000-M3 ASPERGILLUS FUMIGATUS	963	2024-10	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR

6	2	2	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*



CONTROL SERUM

6	2	2	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Blue*

DATE 03/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.