



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-FI EGG WHITE**

Code	88000- F1		
Lot No.	963	2024-08	2023-10

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3381	88000-FI EGG WHITE	963	2024-08	2x6
B3380	CALIBRATORE 88000-FI EGG WHITE	963	2024-10	1
C3380	CONTROL SERUM 88000-FI EGG WHITE	963	2024-10	1

**CONTROLS**

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

**CALIBRATOR**

6	1	1	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

611963651662248700  
Egg white  
CALIBRATOR 0.250 mL [IVD]  
LOT 963  
2024-10 2°C - 8°C  
DIESSSE Monteriggioni Italy CE

**CONTROL SERUM**

6	1	1	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

611963630550000001  
Egg white  
CONTROL 0.450 mL [IVD]  
LOT 963  
2024-10 2°C - 8°C  
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 03/10/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.