

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS BORRELIA IgG CSF

Code	81048		
Lot No.	965	2024-08	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D9620	BORRELIA IgG CSF	965	2024-08	2X6
B9620	CALIBRATORE BORRELIA IgG CSF	965	2024-09	1
C9620	CONTROL BORRELIA IgG CSF	965	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	390	≥ 100
Control Serum (Index)	3,2	2,4-4,4

CALIBRATOR

9	1	5	9	6	5	6	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature *Alberici*



CONTROL SERUM

9	1	5	9	6	5	6	2	4	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *Alberici*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alberici*

DATE *04-10-2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.