

DIESE DIAGNOSTICA SENESE SpA  
 53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
 20144 Milano (Italy) Via A. Solari 19 sc. 6

**PRODUCT NAME:** VACU-CODE

<b>Code</b>	10230		
<b>Lot No.</b>	3923	2024-08	2023-09
<b>Reference</b>	IO-09-88		

**ON LINE QUALITY CONTROLS** (model CDP not enclosed)

**QC Passed**

- sodium citrate solution concentration
- sodium citrate volume
- vacuum level (\*)
- correct placement of the cap
- no extra body inside the product
- (\*) Only for VACU-CODE and VACU-TEC tubes

**PACKAGING QUALITY CONTROL** (model CDP not enclosed)

**QC Passed**

- control of the number of the product for each pack

**QUALITY CONTROLS ON FINISHED PRODUCT** (model CDP not enclosed)

**QC Passed**

- pack integrity
- data printed on the product label

**QC Passed**

N/A

**Irradiation treatment Certificate n° ..... date ....**

**QC Passed**

N/A

**Gamma rays treatment process in compliance with the specifications of the supply Agreement**

Document number:

ODV N.A.

***This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.***

**APPROVED**

RESPONSIBLE HEAD, Quality Control Reagents

**DATE 03/10/23**