



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgM-CHORUS

| | | | |
|---------|-------|---------|---------|
| Code | 81011 | | |
| Lot No. | 963 | 2024-11 | 2023-08 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|-----------------------------------|------|---------|-----|
| D2171 | Cytomegalovirus IgM | 963 | 2024-11 | 6X6 |
| B2171 | CALIBRATORE Cytomegalovirus IgM | 963 | 2024-11 | 1 |
| C2171 | CONTROL SERUM Cytomegalovirus IgM | 963 | 2024-11 | 1 |
| G2171 | CONFIRMATORY TEST | 963 | 2024-11 | 5X1 |

CONTROLS

| | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

| | | |
|--------------------------|-----|---------|
| Calibrator (mOD) | 550 | ≥ 100 |
| Control Serum (INDEX) | 1,8 | 1,4-2,6 |
| Negative Control (INDEX) | 0,0 | <0,7 |

079963700000510000
Cytomegalovirus IgM
CALIBRATOR 0.175 mL IVD
LOT 963
2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE 0123 B2171

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 7 | 9 | 9 | 6 | 3 | 7 | 0 | 0 | 0 | 0 | 0 | 5 | 1 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature HB

CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 7 | 9 | 9 | 6 | 3 | 7 | 1 | 4 | 2 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature HB

079963714260000001
Cytomegalovirus IgM
CONTROL+ 0.425 mL IVD
LOT 963
2024-11 2°C-8°C
DIESSE Monteriggioni Italy CE 0123 C2171

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... HB

DATE 10/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!