

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** DEAMIDATED GLIADIN-G - CHORUS

Code	86100		
Lot No.	972	2024-09	2023-09

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	972	2024-09	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	972	2024-09	1
C8660	CONTROL SERUM Deamidated Gliadin-G	972	2024-09	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	450	≥ 100
Control Serum (AU/mL)	137,7	55,6-194,7

**CALIBRATOR**

2	5	5	9	7	2	6	1	1	8	2	7	3	6	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

255972611827364600  
DEAMIDATED GLIADIN-G  
CALIBRATOR 0.175 mL [TVB]  
LOT 972  
 2024-09 2°C-8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

2	5	5	9	7	2	6	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

255972619360000001  
DEAMIDATED GLIADIN-G  
CONTROL 0.425 mL [TVB]  
LOT 972  
 2024-09 2°C-8°C  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 04/10/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.