



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgA-CHORUS

Code	81062		
Lot No.	963	2024-08	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2401	Helicobacter Pylori IgA	963	2024-08	6X6
B2401	CALIBRATORE Helicobacter Pylori IgA	963	2024-08	1
C2401	CONTROL SERUM Helicobacter Pylori IgA	963	2024-08	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	470	≥ 100
Control Serum (INDEX)	3,1	2,4-4,4
Negative Control (INDEX)	0,0	<0,7

CALIBRATOR

0	2	7	9	6	3	6	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



027963600000300000
Helicobacter pylori IgA

CALIBRATOR 0.175 mL IVD

LOT 963

2024-08 2°C-8°C

DIESSSE Monteriggioni Italy

B2401

CONTROL SERUM

0	2	7	9	6	3	6	2	4	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



027963624440000001
Helicobacter pylori IgA

CONTROL+ 0.425 mL IVD

LOT 963

2024-08 2°C-8°C

DIESSSE Monteriggioni Italy

C2401

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagent

DATE *16/10/23*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.