



DIESSE DIAGNOSTICA SENESE Spa
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TETANUS IgG - CHORUS

Code 81270

Lot No. 971 2024-09 2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D9660	Tetanus IgG	971	2024-09	6X6
B9660	CALIBRATORE Tetanus IgG	971	2024-09	1
C9660	CONTROL SERUM Tetanus IgG	971	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	517	≥ 100
Control (mIU/mL)	448,7	350,0-708,3

CALIBRATOR

9 0 2 9 7 1 6 2 2 0 3 2 5 2 5 1 0 0

Label checked with barcode reader YES Operator Signature *[Signature]*

902971622032525100
Tetanus IgG
CALIBRATOR 0,176 mL [IVD]
LOT 971
2024-09 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9 0 2 9 7 1 6 1 4 2 5 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *[Signature]*

902971614250000001
Tetanus IgG
CONTROL 0,426 mL [IVD]
LOT 971
2024-09 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 05.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.