

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-M1 PENINCILLIUM NOTATUM

Code	88000-M1		
Lot No.	963	2024-08	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D1151	88000-M1 PENINCILLIUM NOTATUM	963	2024-08	2X6
B1150	CALIBRATORE 88000-M1 PENINCILLIUM NOTATUM	963	2024-10	1
C1150	CONTROL SERUM 88000-M1 PENINCILLIUM NOTATUM	963	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR

6	2	8	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*

628963651862248700
Penicillium notatum
CALIBRATOR 0.250 mL TVD
LOT 963
2024-10 2°C 8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	2	8	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*

628963630550000001
Penicillium notatum
CONTROL+ 0.450 mL TVD
LOT 963
2024-10 2°C 8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Blusa*

DATE *12.10.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.