



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-D2 DERMATOPHAGOIDES FARINAE

Code	88000-D2		
Lot No.	963	2024-08	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D9801	88000-D2 DERMATOPHAGOIDES FARINAE	963	2024-08	2X6
B9800	CALIBRATORE 88000-D2 DERMATOPHAGOIDES FARINAE	963	2024-10	1
C9800	CONTROL SERUM 88000-D2 DERMATOPHAGOIDES FARINAE	963	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR

6 0 3 9 6 3 6 5 1 8 6 2 2 4 8 7 0 0

Label checked with barcode reader YES  Operator Signature

603963651862248700  
Dermatophagoides  
farinae  
CALIBRATOR 0.280 mL IVD  
LOT 963  
2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6 0 3 9 6 3 6 3 0 5 5 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature

603963630550000001  
Dermatophagoides  
farinae  
CONTROL+ 0.450 mL IVD  
LOT 963  
2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 17/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.