

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-E2 DOG EPITHELIUM

Code	88000-E2		
Lot No.	963	2024-08	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D1201	88000-E2 DOG EPITHELIUM	963	2024-08	2x6
B1200	CALIBRATORE 88000-E2 DOG EPITHELIUM	963	2024-10	1
C1200	CONTROL SERUM 88000-E2 DOG EPITHELIUM	963	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR

6	1	8	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

618963651862248700
Dog epithelium
CALIBRATOR 0.280 mL [IVD]
LOT 963
 2024-10 2°C B1200
DIESSA Monteriggioni Italy CE

CONTROL SERUM

6	1	8	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

618963630550000001
Dog epithelium
CONTROL 0.450 mL [IVD]
LOT 963
 2024-10 2°C C1200
DIESSA Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 19.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.