

QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgM

Code	81059		
Lot No.	978	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D2382	Epstein-Barr Early Antigen IgM	978	2024-10	6x6
B2382	CALIBRATORE Epstein-Barr Early Antigen IgM	978	2024-10	1
C2382	CONTROL SERUM Epstein-Barr Early Antigen IgM	978	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	440	> 100
Control Serum (Index)	2,8	2,2-4,1
Negative Control (Index)	0,0	<0,7

CALIBRATOR

0	2	9	9	7	8	6	0	0	0	0	0	3	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



CONTROL SERUM

0	2	9	9	7	8	6	2	2	4	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 20-10-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.