



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F13 PEANUT

Code 88000-F13

Lot No. 963 2024-08

2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D9941	88000-F13 PEANUT	963	2024-08	2x6
B9940	CALIBRATORE 88000-F13 PEANUT	963	2024-10	1
C9940	CONTROL SERUM 88000-F13 PEANUT	963	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR

6 1 7 9 6 3 6 5 1 8 6 2 2 4 8 7 0 0

Label checked with barcode reader YES Operator Signature 88

617963651862248700
Peanut
CALIBRATOR 0.250 mL IVD
LOT 963
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE
B9940

CONTROL SERUM

6 1 7 9 6 3 6 3 0 5 5 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature 88

617963630550000001
Peanut
CONTROL+ 0.450 mL IVD
LOT 963
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE
C9940

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 19.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.