



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-T9 OLEA EUROPEA

Code	88000-T9		
Lot No.	963	2024-08	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D9921	88000-T9 OLEA EUROPEA	963	2024-08	2X6
B9920	CALIBRATORE 88000-T9 OLEA EUROPEA	963	2024-10	1
C9920	CONTROL SERUM 88000-T9 OLEA EUROPEA	963	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	242	≥ 100
Control Serum (KU/L)	12,1	7,5-15,7

CALIBRATOR

6	0	9	9	6	3	6	4	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*

609963641862248700
Olea europea
CALIBRATOR 0.250 mL [IVD]
LOT 963
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6	0	9	9	6	3	6	3	0	5	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *RB*

609963630550000001
Olea europea
CONTROL+ 0.450 mL [IVD]
LOT 963
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *26.10.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.