



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME PR3 - CHORUS

Code	86066			
Lot No.	972	2024-09	2023-09	

REAGENTS

Label	Description	Lot.	Exp.	NO
D3512	PR3	972	2024-09	6X6
B3512	CALIBRATORE PR3	972	2024-09	1
C3512	CONTROL SERUM PR3	972	2024-09	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	369	≥ 100
Control Serum (AU/mL)	70,1	56,0-100,0

CALIBRATOR

2	5	2	9	7	2	6	4	1	4	2	9	2	7	4	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB



CONTROL SERUM

2	5	2	9	7	2	6	2	7	4	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]

DATE 02/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive