



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F48 ONION

Code	88000-F48		
Lot No.	985	2024-10	2023-10

REAGENTS

Label	Description	Lot	Exp.	NO
D1101	88000-F48 ONION	985	2024-10	2x6
B1100	CALIBRATORE 88000-F48 ONION	985	2024-10	1
C1100	CONTROL SERUM 88000-F48 ONION	985	2024-10	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	267	≥ 100
Control Serum (KU/L)	9,5	7,5-15,7

CALIBRATOR

6	8	0	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



CONTROL SERUM

6	8	0	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 02.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.