



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME ASCA-A - CHORUS

Code	86062		
Lot No.	978	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3471	ASCA-A	978	2024-10	6X6
B3471	CALIBRATORE ASCA-A	978	2024-10	1
C3471	CONTROL SERUM ASCA-A	978	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	338	≥ 100
Control Serum (AU/mL)	98,2	64,3-184,6

CALIBRATOR

5	0	7	9	7	8	6	4	1	9	3	3	3	0	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

507978641933305900
ASCA-A
CALIBRATOR 0.175 mL [IVD]
LOT 978
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

5	0	7	9	7	8	6	2	3	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

507978623440000001
ASCA-A
CONTROL 0.425 mL [IVD]
LOT 978
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents FB

DATE 02/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.