



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-F95 PEACH**

Code	88000-F95		
Lot No.	985	2024-10	2023-10

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D5751	88000-F95 PEACH	985	2024-10	2X6
B5750	CALIBRATORE 88000-F95 PEACH	985	2024-10	1
C5750	CONTROL SERUM 88000-F95 PEACH	985	2024-10	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	267	≥ 100
Control Serum (KU/L)	9,5	7,5-15,7

**CALIBRATOR**

6	4	7	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature ..... *RB*

647985651862248700  
Peach  
CALIBRATOR 0.250 mL IVD  
LOT 985  
2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

6	4	7	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature ..... *RB*

647985630550000001  
Peach  
CONTROL+ 0.450 mL IVD  
LOT 985  
2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *02.11.2023*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.