



QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code	86074		
Lot No.	979	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	979	2024-10	6X6
B3591	CALIBRATORE a-TPO	979	2024-10	1
C3591	CONTROL SERUM a-TPO	979	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (IU/mL)	405,3	231,6-666,7

CALIBRATOR

3	0	1	9	7	9	6	6	2	3	4	2	2	8	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*

301979662342285700
a-TPO
CALIBRATOR 0.175 mL IVD
LOT 979
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

3	0	1	9	7	9	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *[Signature]*

301979625470000001
a-TPO
CONTROL 0.425 mL IVD
LOT 979
2024-10 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 06-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.